TEST REPORT
IEC/EN 60601-1-2: 2007 (Third Edition)
Medical Electrical Equipment
PART 1-2: General Requirements for Basic Safety and Essential Performance
Collateral Standard: Electromagnetic Compatibility

Report Reference No.	17-07-MAS-055
Date of issue	2018-05-08
Total number of pages	65
Testing	
Date of receipt of test item	2017-07-17
Date (s) of performance of tests	2017-09-04
Testing Laboratory	ELECTRONICS TESTING CENTER, TAIWAN
Address	No.8, Lane 29, Wenming Rd., Guishan Dist., Taoyuan City 33383, Taiwan, R.O.C.,
E-Mail	Suede@etc.org.tw
Applicant's name	HORN ENTERPRISE CO., LTD.
Address	NO.6, LN. 833, DAXUE W. RD., NANZI DIST., KAOHSIUNG CITY 811, 407. TAIWAN(R.O.C)
Test specification	
Standard	IEC/EN 60601-1-2:2007 Emissions: CISPR 11:2009/A1:2010 Immunity: IEC 61000-4-2:2008; IEC 61000-4-3:2006+A1:2007+A2:2010; IEC 61000-4-8:2009
Test procedure	M00-TM-073
Non-standard test method	N/A
Test item description	HNCare Health Wrist Watch
Trade Mark	*HNCare*
Manufacturer	HORN ENTERPRISE CO., LTD.
Model/Type reference	B7
Series Model	----
Ratings	USB: 5V; Li-ion Battery: 3.7V

Possible test case verdicts:
- test case does not apply to the test object ..... N (N/A)
- test object does meet the requirement ..... P (Pass)
- test object does not meet the requirement ..... F (Fail)
- test object does not evaluation the requirement ..... N/E

General remarks:
The test results presented in this report relate only to the object tested.
The results contained in this report reflect the results for this particular model and serial number.
It is the responsibility of the manufacturer to ensure that all production models meet the intent of the requirements detailed within this report.
This report shall not be reproduced, except in full, without the written approval of the issuing testing laboratory.
"(see Enclosure #)" refers to additional information appended to the report.
"(see appended table)" refers to a table appended to the report.
Throughout this report the decimal separator is comma or point. (Select one).

Test Engineer	Checked By	Approved By
Ron Lin	Kuang Chi Yu	Jerry Huang